

# CERTIFICATE OF ANALYSIS

**Complement Technology, Inc.**  
4801 Troup Hwy, Suite 701  
Tyler, Texas 75703, USA

Product: **VBS<sup>++</sup>**  
Catalog # **B113**  
Prep. Date: \_\_\_\_\_

<u>Specifications</u>	<u>Limits</u>	<u>Results</u>
FILL VOLUME	1000 – 1100 mL	1050 mL
BUFFER	Veronal Buffered saline plus calcium and magnesium, pH 7.3 ± 0.1	Veronal Buffered saline plus calcium and magnesium, pH 7.3 ± 0.1
PRESERVATIVE	Sodium Azide, 0.025%	Sodium Azide, 0.025%
FILTRATION	0.22 µm filter	0.22 µm filter
IONIC STRENGTH	14 mmho ± 1 @ 22 °C	14 mmho ± 1 @ 22 °C

## STORAGE CONDITIONS

Buffer may be stored for 3 months at 4 °C.  
Buffer may be stored for 4 years at -20 °C. If frozen, buffer should be warmed and stirred prior to use. Buffer should then be chilled to 4 °C before use.

FOR RESEARCH USE ONLY  
NOT FOR HUMAN OR DRUG USE

---

Signature of Analyst

---

Date of Analysis