

# CERTIFICATE OF ANALYSIS

**Complement Technology, Inc.**  
4801 Troup Hwy, Suite 701  
Tyler, Texas 75703, USA

Product: **Factor I**  
Catalog # **A138 Lot #**  
Exp. Date:

Description: Factor I Purified Human Complement Protein

<u>Specifications</u>	<u>Limits</u>	<u>Results</u>
PROTEIN CONCENTRATION	0.95 – 1.2 mg/mL. Factor I has an extinction coefficient of $E^{1\%}_{280\text{nm}} = 14.3$	1.04 mg/ml
FILL VOLUME	0.100 – 0.120 mL	0.110 mL
BUFFER	Phosphate, buffered Saline, pH 7.2	Phosphate, buffered saline, pH 7.2
PRESERVATIVE	None, filtered through a 0.22 $\mu\text{m}$ pore size filter.	None, filtered through a 0.22 $\mu\text{m}$ pore size filter.
PURITY	> 90%	$\geq$ 95% by SDS-PAGE
IMMUNOCHEMISTRY	$\leq$ trace amounts of IgG, IgA, IgM, Factor H, Factor B, Factor P, C3, C4, or albumin	Conforms
HUMAN SERUM/PLASMA STARTING MATERIAL		
HBsAg	Negative	Negative
Anti-HIV 1 and 2	Negative	Negative
Anti-HCV	Negative	Negative
HIV-1 by NAT	Negative	Negative
HTLV I & II	Negative	Negative
STS	Negative	Negative

**STORE AT -70°C or BELOW.**  
**Thaw quickly at 37°C, mix, and put in an ice+water bath to cool.**  
**Avoid Repeated Freeze/Thaw**

**FOR RESEARCH USE ONLY**  
**NOT FOR HUMAN OR DRUG USE**

---

Signature of Analyst

---

Date of Analysis